> | L

# **PiloTREK**

## W-100

2-wire non-contact microwave level transmitter

User's and Programming manual 8<sup>th</sup> edition



WE'VE GOT

Manufacturer: NIVELCO Process Control Co. H-1043 Budapest, Dugonics v. 11. Tel.: (36-1) 889-0100 Fax: 889-0200 E-mail: sales@nivelco.com www.nivelco.com

	Reference document number	
C FM APPROVED	FM Canada, Certificate No.: FM17CA0074X	wes1404a0600s_01
FM	FM US, Certificate No.: FM17US0134X	wes1404a0600s_01
<mark>€x</mark> >	BKI ATEX, Certificate No.: BKI13ATEX0017X/3	wes1404m0600p 07
IEC.	BKI IECEx, Certificate No.: IECEx BKI 13.0005X Issue No.: 2	wes1404a0600p_07
EAC	Ex Russia, Certificate No.: RU C-HU.MF62.B.04401	wes1404o0600q_01
	INMETRO, Certificate No.: DNV 15.0065 X/1	wes1404p0600p_07
FC	Certificate No.: S7W-WES100	



#### **BASIC CONCEPT OF MEASUREMENT WITH PIIOTREK**

### TABLE OF CONTENTS

1.	INTRODUCTION					
2.	ORDER CODES					
3.	TECH	HNICAL DATA	8			
	3.1.	EXPLOSION PROTECTION. EX MARKINGS. EX LIMIT DATA	9			
	••••	311 ATEX APPROVAL No : BK/13ATEX0017X/2	ç			
		3.1.2 IECEX APPROVAL No : IECEX BKI 13.0005X ISSUE No : 1	9			
		3.1.3. FM US APPROVAL No.: FM17US0134X	10			
		3.1.4. FM CANADA APPROVAL No.: FM17CA0074X	10			
		3.1.5. INMETRO APPROVAL No.:DNV 15.0065 X/1	11			
	3.2.	DIMENSIONS AND SPECIAL DATA OF THE ANTENNA VARIATIONS	12			
		3.2.1. DETERMINE THE MAXIMAL MEASURING RANGE	18			
	3.3.	Accessories	19			
	3.4.	Conditions of Safe Operation	19			
	3.5.	REPAIR. MAINTENANCE AND STORAGE CONDITIONS	19			
4	INST	ALLATION	20			
	4 1	Multing	20			
	12	Monte				
	4.2.	WINNO				
		4.2.1. WINING OF THE DEVICES	23			
	43	I OD CIURDENT CHECKING WITH HAND INSTRUMENT	25			
5	PRO	GRAMMING	25			
υ.	5 1		26			
	J.1.		20			
		5.1.1 I RIMARI MEASOREMENT SCREEM.				
		513 Echo Map				
	52	PROGRAMMING WITH THE SAP-300 DISPLAY MODULE	30			
	0121	5.2.1. COMPONENTS OF THE PROGRAMMING INTERFACE	30			
		5.2.2 Menu structure	31			
	5.3.	Programmable Features Description	32			
		5.3.1. BASIC MEASUREMENT SETTINGS	32			
		5.3.2 ANALOGUE OUTPUT	33			
		5.3.3. DIGITAL OUTPUT	35			
		5.3.4. MEASUREMENT OPTIMIZATION	35			
		5.3.5. CALCULATIONS	37			
		5.3.6. Service functions	40			
6.	ERR	OR CODES	42			
7.	PiloT	IREK W-100 PARAMETER TABLE	43			
8.	MEN		45			



#### Thank you for choosing a NIVELCO instrument. We are sure that you will be satisfied throughout its use!

## 1. INTRODUCTION

#### Application

The **PiloTREK W-100** non-contact microwave level transmitters provide the most advanced, new generation measurement technique of the industrial process automation field. **PiloTREK** is an ideal solution of high precision level transmitting of liquids, slurries, dollops, emulsions and other chemicals in a wide range of application area, such as food industry, energy industry, pharmaceutical industry, chemical industry, and even in naval applications with mm accuracy range and high measuring stability.

**PiloTREK** is able to provide an excellent non-contact measurement solution for those substances which tend to steam, or for liquids with a gas layer. Since there is no need for a defined propagation medium in the case of microwaves, the **PiloTREK** is applicable in vacuum.

#### **Operation principle**

The reflection of the emitted microwave impulses is considerably depending on the relative dielectric constant of the measured medium. The essential condition of microwave level measurement is that the relative dielectric constant ( $\epsilon_r$ ) of the medium should be more than 1.9.

The operation of the non-contact microwave level transmitters is based on the measurement of the time of flight of the reflected signals, so-called Time Domain Reflectometry (TDR) method.

The propagation speed of microwave impulses is practically the same in air, gases and in vacuum, independently from the process temperature and pressure, so the measured distance is not affected by the physical parameters of medium to be measured.

The **PiloTREK** level transmitter is a Pulse Burst Radar operating at 25 GHz (K-band) microwave frequency.

The 25 GHz models' most noticeable advantage over the lower frequency (5 – 12 GHz) radars are the smaller antenna size, the better focusing, lower dead-band and smaller transmission angle.

The level transmitter induces few nanosecond length microwave impulses in the antenna and a part of the energy of the emitted signals reflects back from the measurement surface depending on the measured media. The time of flight of the reflected signal is measured and processed by the electronics, and then this is converted to distance, level or volume proportional data.

## 2. ORDER CODES (NOT ALL COMBINATIONS POSSIBLE!)

PiloTREK	W
----------	---

_			_		_
	-	1		_	
					<u> </u>

FUNCTION	CODE
2-wire compact	E
2-wire compact + display	G
2-wire integrated	Р
High temperature	Н
High temperature transmitter + display	J

ANTENNA / HOUSING MATERIAL	CODE
1.4571 / Aluminium housing	S
1.4571 / Plastic housing	М
PP / Plastic housing	Р
1.4571 / Stainless steel housing	к

ANTENNA TYPE / PROCESS CONNECTION	CODE
Parabola DN150 / Flange	1
DN40 Horn / 11/2"	4
DN50 Horn / 2"	5
DN80 Horn / Flange	8
Planar / 2" (10 m)*	Α
Planar / 2" (16 m)*	В
Egg / 1"	D
Egg / 11/2"	E
Egg / 2"	F

PROCESS CONNECTION	CODE
BSP	0
NPT	N
DN80 PN25	2
DN100 PN25	3
DN125 PN25	4
DN150 PN25	5
DN80, PP	6
DN100, PP	7
DN125, PP	8
DN150, PP	9
3" RF 150 psi	Α
4" RF 150 psi	В
5" RF 150 psi	C
6" RF 150 psi	D
3" FF, PP	E
4" RF, PP	F
5" RF, PP	G
6" RF, PP	Н
JIS 10K80A	J
JIS 10K100A	ĸ
JIS 10K125A	L
JIS 10K 150A	М
JIS 80A, PP	Р
JIS100A, PP	R
JIS125A, PP	S
IIS 150A PP	Т

ОИТРИТ / ЕХ	Co	DDE
4 – 20 mA + HART		4
4 – 20 mA + HART / Ex ia IIIC (ATEX, IECEx)		5
4 – 20 mA + HART / Ex ta/tb IIIC (ATEX,IECEx)		6
4 – 20 mA + HART / Ex ia IIB (ATEX, IECEx)		8
4 – 20 mA + HART /XP IS Div 1 (FM) / ½" NPT		A
4 – 20 mA + HART / NI Div 2 (FM) / ½" NPT		В
4 – 20 mA + HART / Ex db [ia Ga] (ATEX)		с

#### Notes:

S,M,K: 1.4571 stainless steel horn antenna without housing.

P: Aluminium hom antenna with antistatic plastic enclosure, only with DN40 and DN50 hom antenna and thread connection. High temperature version only with metal housing! Antenna seal: Viton. Planar antenna only in integrated type!

\*only for WPP planar antenna type

		• • • • • •
ACCESSORIES TO BE ORDERED NON-SEPARATELY	ORDER CODES	4" RF, PP
PP antenna enclosure with 11/2" BSP threaded process connection / Ex type	WAP-140-0 / WAP-140-8	5" RF, PP
PP antenna enclosure with 11/2" NPT threaded process connection / Ex type	WAP-14N-0 / WAP-14N-8	6" RF, PP
PTFE antenna enclosure with 11/2" BSP threaded process connection	WAT-140-0	JIS 10K80
PTFE antenna enclosure with 11/2" NPT threaded process connection	WAT-14N-0	JIS 10K10
PP antenna enclosure with 2" BSP threaded process connection / Ex type	WAP-150-0 / WAP-150-8	JIS 10K12
PP antenna enclosure with 2" NPT threaded process connection / Ex type	WAP-15N-0 / WAP-15N-8	JIS 10K 1
PTFE antenna enclosure with 2" BSP threaded process connection	WAT-150-0	JIS 80A, F
PTFE antenna enclosure with 2" NPT threaded process connection	WAT-15N-0	JIS100A,
PTFE antenna enclosure with 2" TriClamp process connection	WAT-14T-0	JIS125A,
PTFE antenna enclosure with DN50 Pipe coupling process connection	WAT-14R-0	JIS 150A,

## 3. TECHNICAL DATA

ТҮРЕ		PLASTIC HOUSING WOM-100-0, WOP-100-0	METAL HOUSING WOS-100-0, WOK-100-0	HIGH TEMPERATURE VERSION WHD-100-0, WJD-100-0		
Measured media, measured and calculated values		Liquids; Level, distance, volume, mass				
Frequency of the measuring signal			~25 GHz (K-band)			
Minimal and	I maximal measuring range*					
Material of	wetted parts					
Process co	nnection		See chapter 3.2			
Beam angle	9					
Minimal ε <sub>r</sub> c	f the medium*					
Maximal me	edium pressure (depending on the antenna)	3 bar [43.5 psi] at 25 °C [77 °F]	25 bar [362.6 psi]	at 120 °C [248 °F]		
Medium ter	nperature**	-30 ℃ +100 ℃ [-22 ℉ +21 with PP antenna	2 °F] (up to max. 2 min.: 120 °C [77 °F]), enclosure: 80 °C [176 °F]	-30 °C +180 °C [-22 +356 °F]		
Ambient ter	nperature	-20 °C +60 °C [-4 °F+140 °F]				
Resolution		1 mm (0.04 inch)				
Typical line	arity error (as per MSZ EN 61298-2)*	<0.5 m: ±25 mm, 0.5 – 1m: ±15 mm, 1 – 1.5 m: ±10 mm, 1.5 – 8 m: ±3 mm, >8 m: ±0.04% of the measured distance [< 1.6 ft: ±1 inch, 1.6 – 3.3 ft: ±0.6 inch, 3.3 – 4.9 ft: ±0.4 inch, 4.9 – 26 ft: ±0.12 inch, >26 ft: ±0.04% of the measured]				
Temperatu	e error (as per MSZ EN 61298-3)	0.05% FSK / 10 °C (-20 +60 °C [-4 +140 °F])				
	Analogue	4 – 20 mA (3.95 – 20.5 mA)				
Output	Digital communication	HART <sup>®</sup> (minimal terminal resistor: 250 Ohm)				
	Display	SAP-300 graphical display unit				
Damping tir	ne	Selectable: 0 – 99 sec				
Measuring	requency	10 – 60 sec as per the application settings				
Error indica	tion	Output = 22 mA or 3.8 mA (Ex d[ia]: 3.9 mA)				
Output load		$R_t = (U_t - 20 V) / 0.022 A, U_t =$ power supply voltage				
Power supp	ly voltage	20 V – 36 V DC, Ex ia: 20 V – 30 V DC, Ex d[ia]: 24 V – 36 V DC				
Electrical p	otection	Class III				
Ingress protection		IP67, integrated type (WPM or WPP): IP68				
Electrical connection		M20x1.5 cable glands, cable outer diameter: Ø7 – Ø13 mm (Ø0.28 in – Ø0.51 in), cross section: max. 1.5 mm <sup>2</sup> (15 AWG)				
Housing material		Paint coated aluminium (EN AC 4200) or plastic (PBT)				
Sealing			Viton®, EPDM			
Mass		1 – 1.6 kg	2 – 2.6 kg	2.7 – 3.3 kg		

\*Examined in case of proper application settings at 95% sample rate level. The environment should be free of EMC noises and power supply voltage fluctuations in accordance to the standard, under constant temperature. The reflector should be a plane plate reflector with ideal material, surface and dimensions (min. 3 m x 3 m [10 x 10 ft]). The largest false echo should be 20 dB smaller than the useful echo.

\*\*In case of integrated type transmitters, if the enclosure can be directly in contact with the measured medium, the permissible medium temperature is limited to the ambient temperature.

## 3.1. EXPLOSION PROTECTION, EX MARKINGS, EX LIMIT DATA

#### 3.1.1. ATEX APPROVAL No.: BKI13ATEX0017X/2

ТҮРЕ	PLASTIC HOUSING COMPACT WDM-1DD-D	PLASTIC HOUSING INTEGRATED WPM-1	METAL HOUSING WOS-100-0 WOK-100-0	High Temperature version with Metal housing WHD-1DD-D, WJD-1DD-D
Ex marking (ATEX)	ि II 1/2 G Ex ia IIB T6T5 Ga/Gb	윤 II 1 G Ex ia IIB T6T5 Ga	🖾 ll 1 G Ex ia IIB T6T4 Ga	🕢 ll 1 G Ex ia IIB T6T3 Ga
			ि II 1/2 D Ex ia IIIC T85°CT110°C Da/Db	⟨ II 1/2 D Ex ia IIIC T85°C…T180°C Da/Db
			⟨ II 1/2 D Ex ta/tb IIIC T85°CT110°C Da/Db	⟨ II 1/2 D Ex ta/tb IIIC T85°CT180°C Da/Db
			⟨E͡͡͡͡͡͡͡͡͡ː II 1/2 G Ex db [ia Ga] IIB T6T4 Ga/Gb	🕢 ll 1/2 G Ex db [ia Ga] llB T6T3 Ga/Gb
Ex power supply, Intrinsically safety data	Li: 200µH Ci: 16 nF Ui: 30 V li: 140 mA Pi: 1 W	Li: 200 µH Ci: 30 nF Ui: 30 V li: 140 mA Pi: 1 W	Li: 200 µH Ci: 16 nF Ui: 30 V Ii: 140 m Pi: 1 W "Ex db [ia Ga]": Ut = 24 – 36 V DC, Um=250 V	Li: 200 µH Ci: 16 nF Ui: 30 V Ii: 140 mA Pi: 1 W "Ex db [ia Ga]": Ut = 24 – 36 V DC, Um=250 V

#### 3.1.2. IECEX APPROVAL No.: IECEX BKI 13.0005X ISSUE NO.:1

ТҮРЕ	PLASTIC HOUSING COMPACT WDM-1DD-D	PLASTIC HOUSING INTEGRATED WPM-1	ALUMINIUM HOUSING WOS-100-0 WOK-100-0	High Temperature version with Metal housing WHD-1DD-D, WJD-1DD-D
			Ex ia IIB T6T4 Ga	Ex ia IIB T6T3 Ga
	Ex ia IIB T6…T5  Ga/Gb	Ex ia IIB T6T5 Ga	Ex ia IIIC T85°CT110°C Da/Db	Ex ia IIIC T85°CT180°C Da/Db
			Ex ta/tb IIIC T85°C…T110°C Da/Db	Ex ta/tb IIIC T85°CT180°C Da/Db
			Ex db [ia Ga] IIB T6T4 Ga/Gb	Ex db [ia Ga] IIB T6…T3 Ga/Gb
Ex power supply, Intrinsically safety data	Li: 200 µH Ci: 16 nF Ui: 30 V li: 140 mA Pi:1 W	Li: 200 µH Ci: 30 nF Ui: 30 V li: 140 mA Pi: 1 W	Li: 200 µH Ci: 16 nF Ui: 30 V Ii: 140 mA Pi: 1 W "Ex db [ia Ga]": Ut = 24 – 36 V DC, Um=250 V	Li: 200 µH Ci: 16 nF Ui: 30 V Ii: 140 mA Pi: 1 W "Ex db [ia Ga]": Ut = 24 – 36 V DC, Um=250 V

#### Temperature limit data for ATEX and IECEx approvals:

	HAZARDOUS GAS ATMOSPHERES								EXPLOSIVE DUST ATMOSPHERES			
	PLASTIC H	OUSING	METAL HOUSING					METAL HOUSING				
	WOM-100-0 WOP-100-0					HIGH					Нідн	
TEMPERATURE DATA			WOS-100-0 WOK-100-0		TEMPERATURE WHO-100-0 WJO-100-0		WOS-100-0 WOK-100-0			TEMPERATURE		
										WHD-100-0		
										WJD-100-0		
	Ex ia IIB		Ex ia IIB, Ex db [ia			a Ga] IIB		Ex ia IIIC, Ex ta/tb IIIC		tb IIIC		
Maximum permissible medium temperature	+80°C	+95°C	+80°C	+95°C	+100°C	+130°C	+180°C	+80°C	+95°C	+100°C	+180°C	
Maximum permissible ambient temperature	+60°C											
Maximum resulting surface temperature	+80°C	+95°C	+80°C	+95°C	+100°C	+130°C	+133°C	+80°C	+95°C	+100°C	+133°C	
Temperature class	Т6	T5	T6	T5	Τ4	T4	T3	T85°C	T100°C	T110°C	T180°C	

#### 3.1.3. FM US APPROVAL NO.: FM17US0134X (SEE 'SAFETY MANUAL')

The following data is for information purposes only. The FM certificate and the safety instructions can be found in the attached 'Safety Manual'.

ТҮРЕ	DUAL COMPARTMENT WITH METAL HOUSING WES-100-A, WGS-100-A	HIGH TEMPERATURE VERSION DUAL COMPARTMENT WITH METAL HOUSING WHS-1	High temperature version Dual     Dual compartment     H       compartment with metal housing     with Metal housing     compartment       WHS-1□□-A, WJS-1□□-A     WES-1□□-B, WGS-1□□-B     W				
Marking (FM US)	Class I, Division 1, Group C, I	), T6 Ta = -20°C to +60°C, IP67	Class I, Division 2, Group C, D, T6 Ta = -20°C to +60°C, IP67				
Maximum power supply		35 V DC					
Maximum current	22 mA						
Maximum Voltage		Um:	250 V				

#### 3.1.4. FM CANADA APPROVAL No.: FM17CA0074X (SEE 'SAFETY MANUAL')

The following data is for information purposes only. The FM certificate and the safety instructions can be found in the attached 'Safety Manual'.

TÍPUS	DUAL COMPARTMENT WITH METAL HOUSING WES-100-A, WGS-100-A	HIGH TEMPERATURE VERSION DUAL COMPARTMENT WITH METAL HOUSING WHS-1	DUAL COMPARTMENT         High temperature versity           WITH METAL HOUSING         COMPARTMENT WITH METAL           WES-1□□-B, WGS-1□□-B         WHS-1□□-B, WJS-1□□-B					
Marking (FM US)	Class I, Division 1, Group C, I	), T6 Ta = -20°C to +60°C, IP67	Class I, Division 2, Group C, D, T6 Ta = -20°C to +60°C, IP67					
Maximum power supply		35 V DC						
Maximum current	22 mA							
Maximum Voltage		Um: 250 V						

#### 3.1.5. INMETRO APPROVAL No.: DNV 15.0065 X/1

TYPE	PLASTIC HOUSING COMPACT WDM-1DD-D	METAL HOUSING WDS-100-0 WDK-100-0	HIGH TEMPERATURE VERSION WITH METAL HOUSING WHD-10D-D, WJD-10D-D
		Ex ia IIB T6T3 Ga	Ex ia IIB T6T3 Ga
Ex marking (ATEX) Ex ia IIB T6T5 Ga/Gb	Ex ia IIB T6…T5 Ga/Gb	Ex ia IIIC T85°CT110°C Da/Db	Ex ia IIIC T85°C…T180°C Da/Db
		Ex ta IIIC T85°CT110°C Da/Db	Ex ta IIIC T85°C…T180°C Da/Db
Ex power supply, Intrinsically safety data	Li: 200 µH Ci: 16 nF Ui: 30 V Ii: 140 mA Pi: 1 W	Li: 200 µH Ci: 16 nF Ui: 30 V li: 140 mA Pi: 1 W	Li: 200 µH Ci: 16 nF Ui: 30 V Ii: 140 mA Pi: 1 W

Not all types can be ordered with INMETRO approval.

#### Temperature limit data for INMETRO approval:

		H	AZARDOU	S GAS ATI	EXPLOSIVE DUST ATMOSPHERES					
	PLASTIC HOUSING			Μ	ETAL HOUS	NG	METAL HOUSING			
TEMPERATURE DATA	WOM-100-0 WOP-100-0		WDS-100-0 WDK-100-0		HIGH		WDS.100.0		HIGH TEMPERATURE	
					WHQ-100-0 WJQ-100-0	WOK-100-0			WHD-100-0 WJD-100-0	
				Ex ia IIB				Ex ia	IIIC, Ex ta	a IIIC
Maximum permissible medium temperature	+80°C	+80°C	+80°C	+90°C	+100°C	+180°C	+80°C	+90°C	+100°C	+180°C
Maximum permissible ambient temperature	+60°C									
Maximum resulting surface temperature	+75°C	+80°C	+75°C	+90°C	+100°C	+175°C	+75°C	+90°C	+100°C	+175°C
Temperature class	T6	Т5	Т6	Т5	T4	Т3	T85°C	T100°C	T110°C	T180°C



#### 3.2. DIMENSIONS AND SPECIAL DATA OF THE ANTENNA VARIATIONS











#### 3.2.1. DETERMINE THE MAXIMAL MEASURING RANGE

The maximal possible measuring distance of device types:					
Туре	Antenna type	Measuring distance			
WPP-1A	Planar optoppo	10000 mm (32.8 ft)			
WPP-1B		16000 mm (52.5 ft)			
	DN40 (1 <sup>1</sup> / <sub>2</sub> ") horn antenna with enclosure				
	DN40 (11/2") horn antenna without enclosure	18000 mm (59 ft)			
	TDN50 (2") horn antenna with enclosure	20000 mm (65.6 ft)			
	DN50 (2") horn antenna without enclosure				
WDD-18D	DN80 horn antenna with enclosure	23000 mm (75.5 ft)			
WOO-110	DN150 horn antenna without enclosure				

The maximal measuring distance is illustrated in the diagram on the right in case of materials with different relative dielectric constant. The diagram is valid for horn antenna without plastic enclosure, for liquids with still surface not tending to foaming, vapouring or steaming and in case of ideally slow (<5 m/h [<16.4 ft/hr]) rate of level change.

The maximal measuring range of the **PiloTREK** radars is significantly depending on the circumstances of the application environment and on the selected device type. Depending on the relative dielectric constant of the measuring medium and the process conditions the maximal measurement range (achievable under the reference conditions) may decrease by even 85% (reduce about to one-sixth!). In case of planar antenna version, the maximum measuring distance depending on the dielectric characteristic of the medium to be measured is the same as the encapsulated DN40 horn antenna version.



Depending on the process conditions or the plastic antenna enclosure the following typical reducing factors are recommended to be considered in order to calculate the maximal measuring range. When more than one reducing factors occur at the same time then all the factors should be considered for the calculation:

Process Condition	Reflection reduction in Amplitude	Maximum measuring distance decrease by	Reducing Factor
Slow mixing or slightly waving	2 6 dB	20 50%	08 05
Foaming	2 - 0 dB	20 - 30 %	0.8 - 0.5
Fast mixing, vortex	8 – 10 dB	60 – 70% (the measurement might be completely terminated)	0.4 - 0.3
Steaming, condensation	3 – 10 dB	30 – 70% (the measurement might be completely terminated)	0.7 – 0.3
PP antenna enclosure	2 dB	20%	0.8
PTFE antenna enclosure	1 dB	10%	0.9

For example: Measurement medium is Styrene ( $\epsilon r = 2.4$ ) at 25 °C (77 °F) process temperature and slowly mixed. The device type is WGS-150-4 with WAT-150-0 antenna enclosure. The maximal measuring range is (9 m [29.5 ft] \* 0.5 \* 0.9) = 4 m (13 ft).

18 / 48 • wes1404a0600p\_08

## 3.3. ACCESSORIES

- User's and Programming Manual
- Warrant Card
- EU declaration of Conformity
- 2 pcs. M20x1.5 cable glands
- Sealing (Klinger<sup>®</sup> Oilit) only for BSP threaded process connection

## **3.4. CONDITIONS OF SAFE OPERATION**

To avoid the danger of electrostatic charge accumulation, in case of the **WP**, **WM** types (with plastic electronic housing, plastic antenna enclosure or PP flange) the following safety rule shall be observed:

- The measured medium should be an electrostatic conductor, and the electrical resistivity of the measured medium cannot exceed 10<sup>4</sup> Ω.
- The speed and the method of the filling and emptying process should be chosen properly according to the measured medium.
- The material of the plastic antenna enclosures can produce static electricity. The antenna enclosure is allowed to clean only with wet rag.
- The wiring chamber cover of the Ex d [ia] dual compartment flameproof housing may not be opened while the electrical circuits are energized or if explosive atmosphere is present.

If the **WDs** with aluminium housing is installed into a location which requires 'Ga' protection level, the units should be mounted that they are protected against rare occurring impacts and friction effects which may be source of a potential ignition.

#### Meeting the requirements of the technological process

Please carefully consider that all parts of the instrument which possible to come into contact with the measured medium – including the transducer, the sealing and any other mechanical parts – should meet all requirements of the applied technological process, such as the process pressure, temperature and chemical effects of the used technologies. In addition, it is necessary to take into account at devices with PP flange that the PP flange's mechanical strength is much smaller than the same sized metal flange's.

#### FCC Radio license

This device complies with Part 15 of the FCC Rules. Operation is subject to the following two conditions:

(1) This device may not cause harmful interference, and (2) this device must accept any interference received, including interference that may cause undesired operation. Changes or modifications not expressly approved by the manufacturer could void the user's authority to operate the equipment.

## 3.5. REPAIR, MAINTENANCE AND STORAGE CONDITIONS

PiloTREK units do not require maintenance on a regular basis.

Repair during or after the guarantee period should only be carried out by **NIVELCO**. Devices for repair should be returned fully cleaned, and disinfected. Unused devices must be stored within the ambient temperature range specified in the technical data, and a maximum of 98% relative humidity.

## 4. INSTALLATION

## 4.1. MOUNTING

When choosing the installation place please ensure proper space for later calibrations, verification or maintenance service.

#### PLACEMENT

The ideal position for the **PiloTREK** is on ther = (0.3 - 0.5) R (in case of cylindrical tank).

It is highly recommended to consider the beam cone on the 2<sup>nd</sup> page drawing. The distance between the sensor and the tank wall should be at least 200 mm (7.9 inch). If the unit is installed into dome top or spherical tank, unwanted multiple reflections may appear, which can cancel each other and the measuring signal out, this way it can interfere the measurement.

#### MOVING LIQUID SURFACE

Waving, vortex or strong vibration effects can have negative influence on the measurement accuracy and the maximal measuring range. To avoid these effects, the mounting placement should be as far as possible from the sources of these disturbing effects. According to measurement experiences the maximal measuring distance may decrease by 50 - 70% when the liquid surface is vortexing (see chapter 3.2). For this reason, the device should be mounted as far as possible from the filling stream or the tank outlet.

#### FOAMING

Filling, stirring or any other processes in the tank can generate dense foams on the liquid surface, which may considerably damp the reflected signals.

According to measurement experiences, in these cases the maximal measuring distance decreases at least approximately by 50%.





#### FUMES, VAPOURS

If the measured medium or its foam can reach the antenna or the measured medium is highly fuming, these cases build-ups can form on the sensor, which may result unreliable level measurement.



#### SENSOR ALIGNMENT

The antenna face should be parallel to the medium surface within  $\pm 2^\circ - 3^\circ.$ 



#### TEMPERATURE

To avoid overheating the instrument should be protected against direct sunshine.



M6/SW.3

#### OBSTACLES

Prior to the installation make sure that no objects (cooling pipes, bracing elements, thermometers, etc.) cross the microwave signals. Especially in case of extraordinary large silos bracing elements and other structural obstacles may cause false reflections which can be damped in most cases: a small bent metal deflector plate mounted above the obstacle can disperse the microwave signals and eliminates the false reflections which disturb the reliable measurement.

If there is no possible mechanical solution to avoid these kinds of false reflections, the programming of the instrument allows blocking out the obstacles. (see: 5.3.4.5)

#### EMPTY TANK

Especially in case of standing tanks with hemispherical bottom and in case of tanks which have any equipment inside at the bottom (e.g. heating element, stirrer) wrong level measurement may happen when the tank is totally emptied. The reason for this measurement error is that the tank bottom or the objects at the bottom disperse or reflect the emitted microwave signals. Furthermore, the lower signal-level dispersed radar impulses may interfere with itself inside the tank.

In order to perform reliable level measurement there should be at least 100 mm (4 inch) liquid level above the disturbing objects at the bottom or above the hemispherical tank bottom.



#### POLARIZATION PLANE

The emitted radar impulses of **PiloTREK** are electromagnetic waves. The orientation of the polarization plane is the same as for the electric wave component of the electromagnetic wave. The rotation of the polarization plane compared to the tank position could be useful (for example to avoid disturbing reflections) in certain applications. To rotate the polarization plane loosen the M6 hex socket set screw above the process connection and rotate the instrument. Then tighten the unit by the screw.



The process connection should be implemented that the antenna end should protrude at least 10 mm (0.4 inch) out of the socket.





## 4.2. WIRING

The instrument operates from 20 - 36 V galvanic isolated and not grounded DC power supply in two-wire system. (For Ex version: 20 - 30 V DC!)

The voltage value measured on the terminal of the instrument should be minimum 20 V (in case of 4 mA)! In case of using HART<sup>®</sup> interface – to achieve proper communication between the transmitter's interface and the power-supply – a minimal 250 Ohm resistance should be maintained within the network. The instrument should be wired with shielded cable led through the cable gland. The wiring of the cables can be done after removing the cover of the instrument and the SAP display unit.

**IMPORTANT**: The grounding screw on the housing of the transmitter should be connected to the equipotential network. Resistance of the EP network should be  $R \le 2$  Ohm measured from the neutral point. Shielding of the cable should be grounded at the control room side to the EP network. To avoid disturbing noises, keep away of closeness to high-voltage cables. Especially the inductive couplings of AC harmonics can be critical (which are present at frequency converter control) because even cable shielding does not supply effective protection against these cases.



The instrument may be damaged by electrostatic discharge (ESD) via its terminal, thus apply the precautions commonly used to avoid electrostatic discharge e.g. by touching a properly grounded point before removing the cover of the enclosure. A possible electrostatic discharge can cause damage for the instrument. Do not touch the internal terminals!

#### WATER / VAPOUR

To achieve suitable ingress protection **NIVELCO** recommends using the suggested cable outer diameter (see technical data table in the 3<sup>rd</sup> chapter) and fasten properly the cable gland.

**NIVELCO** also recommends leading the connecting cables downwards to lead aside the rain water and the condensed water. This is needed in case of outside installations and some special applications where there is very high humidity or the possibility of water condensation is quite high (for example in cleaning, purification processes, in cooled and / or heated tanks).



#### 4.2.1. WIRING OF THE DEVICES



#### In case of two-chamber flameproof Ex d [ia] transmitters:

The terminals for the electrical connection of the two-chamber flameproof housing transmitters are in the wiring chamber. The device can be connected to two-wire 4 – 20 mA current loop or HART® multidrop circuit through these terminals. The "+" and "-" wiring terminal points are connected to the "3" and "4" marked pins of the device's electronics through a properly dimensioned - and uncorrectable for the user - certified intrinsically safe and galvanically isolated unit and a certified flameproof conductor. Additional intrinsically safe barriers/circuitry is not required for this device.

#### Extension of the integrated cable:

The usage of a junction box is recommended for extending the cable.

The shielding of the two cables should be connected and grounded at the signal processing device.





MultiCONT

#### In case of integrated version:

Prior to wiring ensure that the power supply is turned off at the source. (For wiring the unit 6 x 0.5 mm<sup>2</sup> cross section or greater cable is recommended). The necessary programming can be made after energizing the unit.



#### 4.2.1. DETERMIN THE APPROPRIATE POWER SUPPLY VOLTAGE

The minimal power supply voltage required by the **PiloTREK** devices is depending on the load impedance in accordance to the below diagram:



In order to provide operation in the total current loop range the calculation should be also checked with 4 mA:

U minimal supply voltage (4 mA) = 4 mA x load resistance + U input minimum (4 mA)

U minimal supply voltage (4 mA) = 4 mA x 250 Ohm + 19 V = 1 V + 19 V = 20 V

Therefore in case of 250 Ohm load resistance 20 V power supply voltage is just enough for the total 4 – 20 mA measuring range.

## 4.3. LOOP CURRENT CHECKING WITH HAND INSTRUMENT

After removing the cover and the Display Module, the actual loop current can be measured throughout an internal 1 Ohm shunt resistor by connecting a voltmeter (in the range of 200 mV) to the points 2 and 5 indicated on the wiring drawing above (see 4.2.1).

## 5. PROGRAMMING

The **PiloTREK** transmitters can be programmed (basically) with the following two ways:

• Programming with the SAP-300 display unit (see 5.2)

All features of the unit can be accessed and all parameters can be set, such as measurement configuration and optimisation, outputs, dimensions for 11 tanks with different shape, 99-point linearization.

• Programming with MultiCONT process controller or EView2 PC configuration software

The PiloTREK WG and WJ types include the SAP-300 display unit.

The **PiloTREK** transmitters are fully operational without the **SAP-300** display as well, it is only needed for local programming and / or local measurement displaying.

#### FACTORY DEFAULT SETTING

The **PiloTREK W-100** series level transmitters are factory programmed by the following way:

- $\Rightarrow$  Measurement mode: Level (LEV). The displayed value is the measured level.
- $\Rightarrow$  The current output and the bargraph on the right are proportional to the measured level.
- $\Rightarrow$  4 mA and 0% are assigned to zero level.
- $\Rightarrow$  20 mA and 100% are assigned to the maximal level.
- $\Rightarrow$  Error indication by the current output: holding the last value.
- $\Rightarrow$  Level tracking time constant: 15 sec.

The instrument regards the distance (DIST) measured from the antenna end as the basic measurement value. This distance is handled and display in one of the selected dimensions: m, cm, mm, feet, or inch. Since the maximal measurement distance is given (entered in P04) the instrument can calculate the actual level (LEV) value. If the proper mechanical dimensions of the mounting – distance between the sealing and the tank bottom – is known, the measured level values can be more accurate by adding this data. The level values calculated that way are the base for volume (VOL) calculation and the 99-point linearization table (VMT) also uses these values as input data.

## 5.1. THE SAP-300 DISPLAY UNIT

#### 5.1.1. PRIMARY MEASUREMENT SCREEN

The **SAP-300** is a 64 x 128 dot-matrix LCD display which can be plugged into the transmitter. (It is universal – usable in other **NIVELCO** devices as well – provided that the system software supports **SAP-300**.)

#### Warning!

The **SAP-300** module is based on LCD technology, so please make sure it is not exposed to permanent heat or direct sunlight, in order to avoid damage of the display unit. If the instrument cannot be protected against direct sunlight or high temperature that is beyond the standard operating temperature range of the **SAP-300**, please do not leave the SAP display in the instrument.

Measurement displaying with the SAP-300 display unit

#### Elements of the displaying:

- 1. Primary (Measured) Value (PV), in accordance to BASIC SETUP / PV. MODE.
- 2. Calculation mode of Primary Value (PV), in accordance to BASIC SETUP / PV. MODE.
- 3. Type and value of the initial quantity used for calculating the Primary Value (PV):
- in case of Level measurement (LEV) it is Distance (DIST),
- in case of Volume measurement (VOL) it is Level (LEV).

4. Trend direction arrows. The empty triangle shows when the change of the measured value is small, the filled triangle shows large-scale change. If none of the arrows are shown the measured value is constant.

5. Measured PV (Distance Value) in relation to measurement range (Sensor range) displayed in a bargraph.

6. Indication of Primary Value simulation. In this case the display and output show the values of the simulation and not the measured values.

7. Indication of active (Volume / Mass Table - VMT) calculation mode.

During active simulation the critical measurement errors will be displayed to give information to the user.





A, Calculated value of the output current.

After the dimension, the mode of current output is indicated by inverse inscription:

#### Μ

Manual mode (see 5.3.2.1)

#### Η

HART address is not 0, so output current has become overwritten to 4 mA (see 5.3.2.1)

#### E!

Analogue transmission reacts to a programmed failure condition if an upper or lower fault current is programmed (see 5.3.2.4)

B, Output range (4 – 20 mA) indicated in a bargraph.

The bottom of the bargraph is assigned to 4 mA and the top is assigned to 20 mA.

C, Indication of Menu Lock:

- If key symbol is visible, the unit is protected with a password. When entering the menu, the instrument asks for the correct password (see 5.3.6.1).
- If REM message is visible, the instrument is in remote programming mode and the main menu cannot be accessed.

Errors occurred during the measurement can be seen at the bottom line of the display.

#### 5.1.2. INFORMATION SCREENS

Press 🕑 button to cycle between the main measurement screen and the information display screen:

- General information screen (DEV. INFO) Overall running time (OV. RUN TIME) Run time after power on (RUN TIME) Type of interface (INTERFACE) in the instrument. Type of the instrument (TYPE)
- 2. Sensor information screen: (SENSOR INFO) Number of echoes (ECHO TOT/SEL) Blocking (BLOCKING) Signal-to-noise ratio (SN) Temperature (TEMP)
- 3. Echo table: (ECHO TABLE)

The location (distance) and the amplitude of the echoes (Dist. / Amp.) are listed

The listed items are the reflections detected by the **PiloTREK** (measured in dB) and the approximate distance from the process connection. The listed values are not accurate measurement values, since around the selected echo (measurement window) there are further measurements and signal processing procedures in order to provide accurate measurement display and level transmission.

The informative screen returns back to main screen after 30 seconds. By pressing the (\*) button the user can return to the main screen any time. Pressing the (E) button in any of the screens the user can enter the main menu. After exiting the menu always the main screen will be shown.



#### 5.1.3. Есно Мар

Pressing the  $\textcircled{\bullet}$  button in the measurement screen the echo map screen will appear. This screen shows the following information:

- 1. Echo diagram
- 2. Actual measured distance
- 3. Maximal measuring range

The echo map screen returns back to main screen after 30 seconds.

By pressing the E button the user can return to the main screen any time. Pressing the E button in any of the screens the user can enter main menu. After exiting the menu always the main screen will be shown.



## 5.2. PROGRAMMING WITH THE SAP-300 DISPLAY MODULE

When entering the menu, the instrument makes a copy of the actual parameters and all changes are done to this duplicated parameter set. During programming the instrument keeps measuring and transmitting with the current (and intact) parameter set. After exiting the menu the instrument replaces the original parameters with the new parameter set and will measure according to the new parameters. This means that the change of the parameters does not become immediately effective when pressing the (E) button!

Entering the menu can be done by pressing the <sup>(E)</sup> button, while exiting the menu can be done by pressing the <sup>(C)</sup> button.

If the instrument is left in programming mode after 30 minutes it will automatically return to measuring mode. If the **SAP-300** display is removed during programming the instrument immediately returns to measuring mode.

As programming with **SAP-300** (manual programming) and HART<sup>®</sup> (remote mode) programming is not possible at the same time, only one programming method could be chosen. Measured values can be read out through HART<sup>®</sup> at any time.

#### 5.2.1. COMPONENTS OF THE PROGRAMMING INTERFACE

The parameters of the instrument are grouped according to their functions. The programming interface consists of lists, dialog windows, edit windows and report windows.

#### Lists

Navigation between the lines of a list can be done by pressing the  $\textcircled{\bullet}$  /  $\textcircled{\bullet}$  buttons. Pressing the E button activates a list item. The selected list item is marked with inverse colour. Exit from a list by pressing the  $\textcircled{\bullet}$  button.

#### Menu list

The Menu list is a specialized list. Its characteristic is that upon selecting a list item we directly get into another list, and these lists are opening from each other in different levels.

The menu header (1) helps to navigate.

Entering the main menu can be done by pressing the  $^{\textcircled{E}}$  button. Navigation between the menu items can be done by pressing the  $^{\textcircled{E}}$  /  $^{\textcircled{E}}$  buttons. Enter to the selected menu by pressing the  $^{\textcircled{E}}$  button. The selected list item is marked with inverse colour.

Exit from a submenu by pressing the  $\textcircled{\bullet}$  button. Pressing the  $\textcircled{\bullet}$  button in the main menu will quit from the programming mode and the instrument will return to measuring mode.



#### **Dialog window**

During the programming the system sends messages or warnings to the users by dialog windows. These usually can be acknowledged by pressing the  $\textcircled{\bullet}$  button or the user can choose between two options (usually YES or NO) by pressing the  $\textcircled{\bullet}$  /  $\textcircled{\bullet}$  buttons. In some cases, one of the parameters has to be changed to correct an error.

#### Edit window

An edit window is used for modifying a numeric parameter value. The selected character can be changed by pressing the / buttons. The cursor can be moved to left, by pressing the button.

The direction of the cursor movement through the digits is right to left. The changed value can be validated by pressing the  $^{(E)}$  button. The software checks if the entered value is appropriate, exiting from the edit window is only possible after entering a correct value. If the entered value is uninterpretable the software sends an error message in the bottom line (1) of the display.

#### Edit window - button combinations

In the edit window the following button combinations are available:

- 1. Recalling the parameters to the state before editing (+ + +, pressed for 3 sec.)
- 2. Recalling default parameters ( ← + , pressed for 3 sec.)

#### 5.2.2. MENU STRUCTURE

#### Main menu

BASIC SETUP	Parameter group of the basic measurement parameters
OUTPUT SETUP	Parameter group of the output parameters
OPTIMIZATION	Parameter group for measurement optimization settings
CALCULATION	Calculations
SERVICE	Service functions, calibration, test and simulation



## 5.3. PROGRAMMABLE FEATURES DESCRIPTION

#### 5.3.1. BASIC MEASUREMENT SETTINGS

#### 5.3.1.1 Default unit system

Parameter:	P00: c, where c: 0, 1.
Path:	BASIC SETUP / UNITS / ENGINEERING SYSTEM
Description:	This should be configured as the first step of the programming
	Here you can choose the default unit system:

- EU European unit system
- US Anglo-Saxon / American unit system

#### 5.3.1.2 Dimension of the default unit system

Parameter: P00: b, and P02: b, or P02: c Path: BASIC SETUP / UNITS / ENO

Description:

BASIC SETUP / UNITS / ENGINEERING UNITS

The dimension of the selected default unit system can be specified in this menu. The selected measurement mode here will define the primary measured value and the displayed value, furthermore it will be the source for the current output:

- BASIC UNITS (m, cm, mm, ft, inch)
- VOLUME / FLOW UNITS (m<sup>3</sup>, I, ft<sup>3</sup>, gallon)
- MASS UNITS (t, t)
- TOT UNITS (m<sup>3</sup>, l, ft<sup>3</sup>, gallon)
- TOT UNITS (sec, min, hour, day)

If the dimension is modified, the device resets all the parameters after a warning message.

#### 5.3.1.3 Primary Value Mode

 Parameter:
 P01: b a
 Default

 Path:
 BASIC SETUP / PV MODE
 Default

 Description:
 This mode determines the primary value and the displayed value. It also determines the value which will be proportional to the output current.
 Default

- DISTANCE
- LEVEL
- VOLUME
- MASS

Default value: mm, m<sup>3</sup>, t (for USA version: inch, ft<sup>3</sup>, t)

FU

Default value:

(for USA version: US)

Default value: L

LEVEL

#### 5.3.1.4 Maximal Measuring Distance

Parameter: Path: Description:	P04 BASIC SETUP / MAX. MEAS.DIST This parameter should be entered all the cases, except distance measurement mode. But it is suggested to be programmed in case of distance measurements in order to avoid the disturbing effects of possible unwanted multiple reflections!	Default value:	
5.3.1.5 Damping Ti Parameter: Path	I <b>ME</b> P20 BASIC SETUR / DAMPING TIME	Default value:	15 sec
Description:	Damping time is used to damp the unwanted fluctuations of the output and display. If the measured value changes rapidly the new value will settle with 1% accuracy after this set time. (Damping is according to the exponential function).		
5.3.1.6 Demo Mode	9		
Parameter: Path: Description:	<ul> <li>P00: d BASIC SETUP / DEMO MODE <ul> <li>OFF: The operation is performed with considering all the application parameters (such as filling, emptying speed, echo selection, etc.)</li> <li>ON: This fast operation mode ignores the application parameters. The demo mode uses a fast algorithm evaluation independently from P25, P26 and P27 parameters. The measurement accuracy and reliable operation between process environments are not guaranteed!</li> </ul></li></ul>	Default value:	OFF
5.3.2. ANALOGUE C	DUTPUT		
Parameter: Path: Description:	<ul> <li>P12: b, where b: 0, 1.</li> <li>OUTPUT SETUP / ANALOG OUTPUT / CURRENT MODE</li> <li>Transmission mode of the current output.</li> <li>AUTO The output current is calculated from the measured value, output is active.</li> <li>MANUAL The output current is fixed at a constant (set) value (see: 5.3.2.5). In this mode the setting of the error current is irrelevant. The set (current) value overwrites the 4 mA output of HART<sup>®</sup> multidrop mode!</li> </ul>	Default value:	AUTO

#### 5.3.2.2 Output Current Value assigned to 4 mA

Parameter: Path: Description:	P10 OUTPUT SETUP / ANALOG OUTPUT / 4 mA VALUE Measured value assigned to 4 mA current value. The transmitted value is in accordance to the primary value (PV) (P01: a). Assignment can be done that the change in measured value and the change in the output value are the same (normal), or opposite directional (inverse operation). For example: 1 m (3.28 ft) level is 4 mA, 10 m (32.8 ft) level is 20 mA, or conversely.	Default value:	0
5.3.2.3 Output	Current Value assigned to 20 mA		
Parameter: Path: Description:	P11 OUTPUT SETUP / ANALOG OUTPUT / 20 mA VALUE Measured value assigned to 20 mA current value. The transmitted value is in accordance to the primary value (PV) (P01: a). Assignment can be done that the change in measured value and the change in the output value are the same (normal), or opposite directional (inverse operation). For example: 1 m (3.28 ft) level is 4 mA, 10 m (32.8 ft) level is 20 mA, or conversely.	Default value: Maximal measurement	range
5.3.2.4 Output	Current Error Mode		
Parameter: Path: Description:	<ul> <li>P12:a, where a: 0, 1, 2</li> <li>OUTPUT SETUP / ANALOG OUTPUT / ERROR MODE</li> <li>Error indication by the current output: <ul> <li>HOLD Error indication has no effect on the output current.</li> <li>LOW Error indication: the output current gets 3.8 mA. (Ex d[ia] type: 3.9 mA)</li> <li>HIGH Error indication: the output current gets 22 mA.</li> </ul> </li> <li>Warning: This error indication is active unless the failure is fixed, or until the failure terminates.</li> </ul>	Default value:	HOLD
5.3.2.5 Fixed (	Dutput Current		
Parameter: Path: Description:	P08 OUTPUT SETUP / ANALOG OUTPUT / MANUAL VALUE Parameter for setting the fixed output current: Values between 3.8 and 20.5 can be entered. The output current will be set to the entered value and analogue transmission will be suspended (see: 5.3.2.1). This error indication overrides all other error indication.	Default value:	4 mA

#### 5.3.3. DIGITAL OUTPUT 5.3.3.1 HART Polling Address Parameter: P19 Default value: 0 OUTPUT SETUP / SERIAL OUTPUT / ADDRESS Path: Description: HART Polling Address The polling address can be set between 0 and 15. For a single instrument the polling address is 0 and the output is 4 - 20 mA (analogue output). If multiple units are used in HART<sup>®</sup> Multidrop mode (max. 15 pcs.) the polling addresses should differ from 0 (1 – 15), in this case the output current will be fixed at 4 mA. 5.3.4. MEASUREMENT OPTIMIZATION 5.3.4.1 Blocking, Dead Zone Parameter: P05 Default value: 300 mm Path: **OPTIMIZATION / DEAD ZONE** (For USA version: 11.8 inch) Description: The instrument ignores all reflections within the dead zone and the close-end blocking distance. The disturbing objects and false reflections which are close to the sensor can be eliminated by entering the dead zone value manually. 5.3.4.2 Echo Selection P25: a, where a: 0, 1, 2, 3 AUTO Parameter: Default value: Path: **OPTIMIZATION / ECHO SELECTION** Description: Selection of Echo within the measuring window. In order to avoid disturbing reflections the instrument forms a so-called measuring window around the reflected signal. The distance measurement is performed with the echo signal within the measurement window. AUTO FIRST • HIGHEST AMPLITUDE • I AST • 5.3.4.3 Emptying Speed Parameter: P27 Default value: 50 m/h Path: **OPTIMIZATION / EMPTYING SPEED** (For USA version: 164 ft/hr) This parameter provides additional protection against echo loss in applications involving Description: very heavy fuming during emptying process. Correct setting increases the reliability of the measurement during the emptying. The parameter must not be smaller than the fastest possible emptying rate of the actual process. It could be maximum 90 m/h (295 ft/h).

#### 5.3.4.4 Filling Speed

Parameter:	P26
Path:	OPTIMIZATION / FILLING SPEED
Description:	This parameter provides additional protection against echo loss in applications involving very heavy fuming during filling process. Correct setting increases the reliability of the measurement during the filling. The parameter must not be smaller than the fastest possible filling rate of the actual technology. It could be maximum 90 m/h (295 ft/h).

#### 5.3.4.5 Background Image

Parar Path:

#### Parameter: OPTIMIZATION / BACKG.ECHO IMAGE / SAVE BACKG. IMAGE

The not-moving disturbing objects inside the tank which generates unwanted false reflections can be blocked out from the measurement range. For this purpose the instrument needs to map the totally empty tank to create a "background image". After this procedure the software will automatically recognise and ignore the reflections coming from the disturbing objects crossing the microwave beam (see 4.1 – Obstacles).

**Warning!** The background image should be saved only when the tank does not contain measurement medium but the disturbing objects inside the tank are not removed. The background image is not recommended to be saved when the tank is filled with the measurement medium since it might result wrong level measurement.

#### 5.3.4.6 Using saved background image

Parameter:	P35: a, where a: 0, 1	Default value:	OFF
Path:	OPTIMIZATION / BACKG.ECHO IMAGE / SAVE BACKG: IMAGE		
Description:	Turning ON or OFF the usage of saved background image during the calculations as per		
	the above 5.3.4.5 point described.		
	<ul> <li>OFF: Ignoring the saved background image.</li> </ul>		
	• ON: Saving background image, damping reflections coming from the disturbing objects.		
5.3.4.7 Thresh	nold value		
Parameter:	P29	Default value:	4 dB
Path:	OPTIMIZATION / TRESHOLD VALUE		
Description:	Defining an upper limit value above the saved background image described in 5.3.4.5 point.		
	The instrument will evaluate the measurement result as a real echo when the reflected		
	signal exceeds the saved background level with the threshold value entered here. Setting		
	the threshold value is useful when the level in the tank and the position of the (small		
	surface) not-moving disturbing object are the same. This case the instrument will not		
	regards the echo signal as false reflection.		

Default value: 50 m/h (For USA version: 164 ft/hr)



#### 5.3.5. CALCULATIONS

5.3.5.1 Specific gravity

 Parameter:
 P32
 Default value:

 Path:
 CALCULATION / SPECIFIC GRAVITY
 Description:

 Description:
 Entering a value (other than "0") of specific gravity in this parameter, the MASS value will be displayed instead of Volume (VOL) in the dimension of tonne or lb/tonne depending on P00 (c) and P02 (b).

#### 5.3.5.2 Volume / Mass Calculation Mode

 Parameter:
 P47: a
 Default value:

 Path:
 CALCULATION / V/M CALC. MODE
 Description:
 Calculation of the volume and mass can be performed with two ways:
 •

 •
 TANK FUNCTION/SHAPE – volume and mass calculation with a tank shape formula. Entering this menu point the table is automatically OFF.
 •

 V/M TABLE – volume and mass calculation with a table. Entering this menu point the table automatically turns ON.

#### 5.3.5.3 Volume / Mass Table

Parameter:

Path: Description:

- CALCULATION / V/M CALC. MODE / V/M TABLE • VIEW/EDIT TABLE
  - ADD ITEM
  - DELETE ITEM

If none of the formulas match perfectly to the characteristics of the needed tank, there is a possibility to use table calculation mode. The device can handle a 99-point table on this purpose and counts values between the neighbouring point pairs with linear interpolation.

The input (left) side of the table contains the level data, the output (right) side contains the volume or mass data. The first point pair of the table should be 0,0. If a long table wanted to be shortened, 0,0 point pair should be entered into the last item of the table and the device modifies the unused point pairs automatically in the background to 0,0. The status (ON or OFF) of the table is shown on a warning message (1) on the bottom line of the display.

All modifications are done on a temporary table. This temporary table becomes valid after exiting.

Modifications during the programming procedure have no effect on the measurement and the transmitting.



0

0

Adding a point pair (ADD ITEM) to the list or pressing (E) button on an existing item, an edit screen will appear. In this edit screen there are two editing filed. Both editing field works as same as editing a parameter. Getting from the first

SHAPE1

field to the second field press the 🖲 button. Pressing 🖲 button in the second field will return back to the previous menu point. When exiting from the last field, the device performs the ordering of the table.

#### Delete item

Edit table:

Moving in the list can be done with (\*) and (\*) buttons, for deleting an item press the (E) button on the selected item. Exiting from the list can be done by pressing the 🕙 button. The table should contain at least 2 items.

SHAPE3

#### 5.3.5.4 Tank Functions / Shape

Parameter:	P40: a, where a: 0,1, 2, 3, 4.	Default value:	0
Path:	CALCULATION / V/M CALC. MODE / TANK FUNCTION / SHAPE		
Description:	<ul> <li>STANDING CYL. – Standing cylindrical tank</li> <li>STD. CYL. CON. BOT. – Standing cylindrical tank with conical bottom</li> <li>STD. RECT. W/CHUTE – Standing rectangular tank with or without chute</li> </ul>		
5.3.5.5 Tank P	SPHERICAL – Spherical tank     Softom Shape		
Parameter:	P40 h where h 0 1 2 3	Default value:	0
Path:	CALCULATION / V/M CALC. MODE / TANK FUNCTION / SHAPE	Boldan value.	Ũ
Description:	This menu only appears, if it has an importance on the selected tank shape type! <ul> <li>SHAPE0</li> <li>SHAPE2</li> </ul>		

Entering the point pairs can be done in arbitrary order, because the device sorts according to ascending order. Both sides of the table have to be strictly monotonic increasing. In case of any error, warning message (see: 6<sup>th</sup> chapter) will appear. When entering again the table an inscription indicates the first wrong line. View table:

0000.0 000000.000 02: 0100.0 000100.000 In VIEW/EDIT TABLE menu point items of the ordered table can be checked. For moving in the list use the 🐑 and 📀 buttons, for editing the selected item use the 🖲 button. Exiting from the list can be done by pressing the 📀 button. 02. VM TABLE ITEM

EDIT/VIEW TABLE



#### 5.3.5.6 Tank Dimensions



0

#### 5.3.6. SERVICE FUNCTIONS

#### 5.3.6.1 Security Codes

#### User codes

Path:	SERVICE / SECURITY / USER LOCK
Description:	Setting or unlocking the user security code.
	The instrument can be protected against unauthorized programming with a 4-digit PIN (Personal Identification Number) code. If either
	of the digits differs from 0 the code is active. If zero is specified, then the secret code has been deleted!
	In case of Active code, this code is requested at menu entry.

#### Service code

Path:	SERVICE / SECURITY / SERVICE LOCK
Description:	Setting of the service code.
	Only for trained personnel!

#### 5.3.6.2 Current Output Test

Parameter:	P80
Path:	SERVICE / OUTPUT TEST / ANALOG OUTPUT / CURRENT VALUE
Description:	Loop current test (mA)
	Entering this Parameter the current value which is proportional to the actual measurement value will appear on the display and the output. In loop current test mode, values between 3.8 and 22 can be entered. The output current will be set to the entered value. The measured current on the output should be equal to the set value. In test mode a dialog window warns the user of the fixed output current until the user exits the warning message window. Exiting can be done by pressing the $(E)$ button.

#### 5.3.6.3 Distance Simulation

This function facilitates the user to be able to check the calculations (tank formula, table), outputs, and the additional processing instruments connected to the output. **PiloTREK** transmitters can perform simulation on the value of a constant or a variable. To start simulation the instrument must return to Measurement mode. In Measurement mode if simulation is in progress, an inverse SIM caption appears on the display.

Simulation mode				
Parameter:	P84: a, where a: 0, 1, 2,	Default value:	OFF	
Path:	SERVICE / DIST SIMUL	ATION / MODE		
Description:	Simulation mode: OFF FIX VALUE MANUAL VALUE TRIANGLE WAVE	No simulation Value of the simulated distance is set according to the lowest va  Value of the simulated distance changes linearly between the lowest and highest values with an adjustable cycle time.	Ilue of the simulation.	t[sec]
	SQUARE WAVE	The simulated value jumps between the lowest and highest values with an adjustable cycle time.		
Simulation cycle				
Parameter: Path: Description:	P85 SERVICE / DIST. SIMUL Cycle time of the simulat	Default value:	60 sec	
Bottom value of the s	simulation			
Parameter: Path: Description:	P86 SERVICE / DIST. SIMUL Lowest value of the simu	ATION / BOTTOM VALUE lation	Default value:	0
Upper value of the sig	mulation			
Parameter: Path: Description:	P87 SERVICE / SIMULATION Highest value of the simu	I / UPPER VALUE Ilation	Default value:	Programmed measurement range
5.3.6.4 Load Defa	ult Values			
Path: Description:	SERVICE / DEFAULTS / This command loads all of After loading the default until the user exits from t confirmation from the user	LOAD DEFAULT default values of the instrument. values the parameters can freely be changed, the effect of the cha he Programming mode and returns to Measurement mode. Before I er because all user parameters will be lost!	nges does not affect on oading the defaults the s	the measurement software asks for a

## 6. ERROR CODES

MESSAGE ON THE SCREEN	ERROR DESCRIPTION	PROCEDURE TO DO	CODE
MEMORY ERROR	Memory error	Contact the service!	1
NO ECHO	Sensor error	Contact the service!	2
EE COM. ERROR	Hardware error (EEPROM communication error)	Contact the service!	3
MATH. OVERLOAD	Calculation overflow	Check the programming!	4
SIGNAL IN N.D.B.	Sensor or calibration error (Measured value is in the close-end dead-zone)	Contact the service!	5
SIGNAL IN F.D.B.	Sensor or calibration error (Measured value is in the far-blocking zone)	Check the installation conditions!	7
VMT SIZE ERROR	Linearization error: Less than two items are in the table.	Check the content of the VMT! See: 5.3.5.3.	12
VMT INPUT ERROR	Linearization table error: monotonicity error in the input (level) side of the table.	Check the content of the VMT! See: 5.3.5.3.	13
VMT OUTPUT ERROR	Linearization table error: monotonicity error in the output (volume or mass) side of the table.	Check the content of the VMT! See: 5.3.5.3.	14
VMT INPUT OV.RNG.	Linearization table error: The measured level is greater than the highest level of the table's input side.	Check the content of the VMT! See: 5.3.5.3. Device performs extrapolation according to the last point pairs!	15
EE CHK ERROR	Parameter checksum error.	Check the programming! For recalculate the checksum modify a parameter and return to Measurement mode. If this error still remains, contact the service!	16
	Parameter integrity error		
INTEGRITY ERROR	(Automatically corrected internal error). Only WARNING message.	Check the programming!	17
AC COM. ERROR	Hardware error	Contact the service!	18
CALIBRATION ERROR	Sensor calibration error	Contact the service!	

#### Functions of information LEDs:

LED state	"VALID" LED "COM" LED		
NOT LIT	there is no appreciable echo	no HART communication	
FLASHING SLOWLY (OR ONE FLASH)	searching for echo within the valid measuring range	flashes while HART communication is in progress	
CONTINUOUSLY LIT	the selected echo is measured by the instrument	programming on the HART line (remote)	
6X SHORT FLASHES TOGETHER	storage EEProm communication error has occurred		
7X SHORT FLASHES TOGETHER	Linearization table (VMT) error		

## 7. PiloTREK W-100 PARAMETER TABLE

Рхх	Parameter name	d	c	b	a
00	Engineering system, dimensions	DEMO mode	Engineering system:	Dimension:	
		0 = Normal mode	0 = EU	(EU) 0 = m, 1 = cm, 2 = mm	
		1 = Demo mode	1 = US	(US) 0 = ft, 1 = inch	
01	Source of Primary Value (PV)				0 = DIST, 1 = LEVEL,
					2 = VOLUME, 3 = MASS
02	Selectable dimensions	<u>TOT:</u>	Time units:	(VOL/F-EU)	Temperature unit:
		(TOT-EU) 0 = m <sup>3</sup> ,	0 = sec	$0 = m^3$ , 1 = liter	0 = °C
		1 = liter	1 = min	(VOL/F-US)	
		$(TOT-US) 0 = ft^3,$	2 = hour	$0 = ft^3$ , $1 = US$ gallon	
		1 = US gallon	3 = day	(MASS-EU)	
				0 = tonne, 1= US tonne	
				(MASS-US)	
				0 = tonne, 1 = lb(pound)	
04	Max. measuring distance	Maximal measuring distance	of the level transmitter can be defi	ined	
05	Blocking / DEAD ZONE	Minimal measuring distance	within the ignores all the measurer	ment values	
08	Fix current output	Fix forced value on the output	it current between 3.8 and 20.5 m/	A for loop current measuring purp	oses (operation mode = manual)
10	4 mA	Measured and transmitted value (PV) assigned to 4 mA current value			
11	20 mA	Measured and transmitted value (PV) assigned to 20 mA current value			
12	Output current mode			Operation mode:	Error indication
				0 = AUTO	of the current output:
				1 = MANUAL	0 = HOLD;
					1 = LOW;
					2 = HIGH
19	HART <sup>®</sup> polling address	HART <sup>®</sup> Short Address of the	level transmitter (0 – 15)		
20	Damping time	Damping time of the accurate transmitted (displayed) value in sec. after a high fluctuation in the measured value (0-999)			
25	Echo selection in the measuring				0 = AUTO;
	window				1 = FIRST
					2 = HIGHEST AMPLITUDE
					3 = LAST
26	Filling speed	Rate of change of the measu	red value (when distance is decrea	asing) which can be just followed	with the level transmitter
27	Emptying speed	Rate of change of the measu	red value (when distance is increa	sing) which can be just followed	with the level transmitter
29	Threshold value	Threshold limit value (0 – 6 dB) for the received echo evaluation			
32	Specific gravity of the medium	Data for mass calculation			

Рхх	Parameter name	d	C	b	a	
35	Background mode				Calculating with the saved background image: 0 = OFF 1 = ON	
40	Tank shape				<ul> <li>0 = Standing cylindrical tank with dome bottom</li> <li>1 = Standing cylindrical tank with conical bottom</li> <li>2 = Standing rectangular tank with or without chute</li> <li>3 = Lying cylindrical tank</li> <li>4 = Spherical tank</li> </ul>	
41-45	Tank dimensions					
47	VMT mode				Operation of the linearization: 0 = OFF, 1 = ON	
60	Overall runtime	Elapsed overall operating he	ours of the level transmitter (w	vorking time) with 0.1 hour ac	curacy. Service data	
61	Runtime after last reset	Elapsed operating hours of	Elapsed operating hours of the level transmitter since the last turning ON with 0.1 hour accuracy. Service data			
70	Number of echoes	Service data				
71	Position of the measuring window	Service data				
74	Signal-to-noise ratio	Service data				
75	Blocking distance value	Service data				
80	Current output test	Fix forced value on the outp	ut current between 3.8 and 20	0.5 mA for checking the accur	acy of the current generator	
84	Simulation				Distance simulation mode:         0 = No simulation         1 = Fix value         2 = Simulation with a manual value: PV = a entered in P86         3 = Simulation between P86 and P87 levels with P85 cycle time (triangle wave)         4 = Simulation between P86 and P87 levels with P85 cycle time (square wave)	
85	Cycle time of DIST simulation	Cycle time of the distance s	imulation in seconds. Default	value: 60 sec		
86	Bottom value of the simulation	Initial value of the distance	simulation in the selected unit	: (e.g.: mm). Default value: 0 (	mm)	
87	Upper value of the simulation	Final value of the distance simulation in the selected unit (e.g.: mm). Its default value is the same as the programmed maximal				
		measurement range.				

## 8. MENU MAP







wes1404a0600p\_08 November 2019 NIVELCO reserves the right to change technical data without notice!